



**ASPEE COLLEGE OF HORTICULTURE AND FORESTRY**

**Navsari Agricultural University  
Navsari - 396 450**



## **QUOTATION**

Sealed quotations are invited for supply of Plate Rotor for dyanamica velocity 18R centrifuge and Autoclave at Navsari Agricultural University, Navsari. The quotations should be sent in favor of **“The Principal, ASPEE College of Horticulture and Forestry, Navsari Agricultural University, Navsari- 396 450, Gujarat” on or before 12<sup>th</sup> August, 2014**. Details of item’s specification along with the terms and conditions may be obtained from NAU website ([www.nau.in](http://www.nau.in)). In case of any discrepancy the decision of Principal, ACHF will be final. The Principal, ACHF reserves the rights to reject any or all the quotations without assigning any reason thereof.

**Principal**

## **QUOTATION FOR BIOTECHNOLOGY DEPARTMENT, ACHF**

**Last date for submission of quotation is 12<sup>th</sup> August, 2014**

Sealed quotations are invited for Supply of the following items as per given specifications:

<b>1. Autoclave</b>
<ul style="list-style-type: none"><li>➤ Fully Automatic Vertical Autoclave</li><li>➤ Unique single lever lock for lid with single hand opening.</li><li>➤ Should have Microprocessor Based Control</li><li>➤ Should be CE certified</li><li>➤ Internal Chamber and lid along with all wetted parts fabricated from Stainless Steel.</li><li>➤ Should have leak proof sealing with Silicone Gasket</li><li>➤ External vertical panels of MS duly heat cured epoxy coated.</li><li>➤ Easy lift and closing of the lid by assisted support.</li><li>➤ Should have Low water detection.</li><li>➤ Should have safety valve for over pressure.</li><li>➤ Should have Pressure interlock on door.</li><li>➤ Should have Castor wheel installed in bottom for easy movement.</li><li>➤ Audio &amp; Visual alarms for sensor open and low water level.</li><li>➤ Should have MCB for electrical safety.</li><li>➤ Independent over temperature safety cut out.</li><li>➤ Working Chamber dimensions should be:- Size (Ø x D 55 x 76 cm.)/ Capacity approximately 180 lit.</li><li>➤ Two additional basket should be supplied with the instrument.</li></ul>
<b>2. Plate Rotor for dyanamica velocity 18R centrifuge</b>
<ul style="list-style-type: none"><li>➤ Plate rotor should be compatible with dyanamica velocity 18R centrifuge</li><li>➤ Micro plate rotor for 96 well format plate</li><li>➤ Maximum speed RPM should be 4800</li></ul>

## **GENERAL TERMS AND CONDITIONS:**

1. The Sealed quotations should be super scribed as **Quotations for Biotechnology Department** at the top of the cover.
2. The quotations should reach to the office of **Dean & Principal, ASPEE College of Horticulture and Forestry, NAU, Navsari -396450 (Gujarat) on or before 12<sup>th</sup> August, 2014 through post or courier only.**
3. The quoted price should be in **Indian Rupees only** inclusive of all taxes and charges FOR delivery at the College Department.
4. The party should submit the **authorization certificate of dealership/manufacturer** for the quoted items.
5. Documentary proof should be submitted for the certificate of incorporation, copy of PAN card and/or latest IT return/sale tax/service tax registration/commission of Indian agent (in case of imported item), ST/CST registration certificate.
6. The color catalogue of quoted model should be preferably attached along with quotation.
7. Comprehensive onsite warranty for 1 year from the date of installation of the items should be offered.
8. The principal, ACHF reserves the right to accept or reject any quotations without assigning reason.
9. In case of any dispute, the decision of Principal, ASPEE College of Horticulture and Forestry, Navsari Agricultural University, Navsari – 396 450, will be final.